

CalCOFI 1311NH: Secondary Oxygen Sensor vs Bottle O2

4sec secondary sensor average vs bottle oxygen (ml/L)

$$y = 1.0362x - 0.0035$$
$$R^2 = 0.9989$$

